

BEAMLINE	SCIENTIFIC TOPIC	ENERGY RANGE <i>keV</i>	BEAM SIZE <i>H x V</i>	NOMINAL FLUX <i>ph/sec</i>	DETECTORS	SAMPLE ENVIRONMENT & <i>Beamline Support Labs</i>	TECHNIQUE
<p>BM26</p> <p><i>DUBBLE (Dual-Belgian Beamlines)</i></p> <p>SCIENTIST IN CHARGE Martin Rosenthal martin.rosenthal@esrf.fr</p>	<p>Biology</p> <p>Chemistry</p> <p>Materials processing</p> <p>Soft Matter</p>	7 - 20	MAX 300 x 300 μm^2		<ul style="list-style-type: none"> ▪ Pilatus 1M detector ▪ Pilatus 300 K detector 	<ul style="list-style-type: none"> ▪ Heating stages (-150 to 600 °C) ▪ In-line Differential Scanning Calorimetry (-150 to 575 °C) ▪ High temperature furnace (max. 1500 °C) ▪ Tensile tester ▪ Shear devices <p>Beamline Support labs</p> <ul style="list-style-type: none"> ▪ Sample preparation lab 	<p>Small Angle X-ray Scattering (SAXS)</p> <p>Wide Angle X-ray Scattering (WAXS)</p>